

Chapter 3.22

Myoelectric Control of Prosthetic Devices for Rehabilitation

Rami N. Khushaba

University of Technology, Australia

Adel A. Al-Jumaily

University of Technology, Australia

INTRODUCTION

Bio-signals patterns analysis problems have enjoyed a rapid increase in popularity in the past few years. The electromyography (EMG) signal, also referred to as the Myoelectric signal (MES), recorded at the surface of the skin, is one of the biosignals generated by the human body, representing a collection of electrical signals from the muscle fibre, acting as a physical variable of interest since it first appeared in the 1940s (Scott, 1984). It was considered to be the main focus of scientists, and was advanced as a natural approach

for the control of prosthesis, since it is utilising the electrical action potential of the residual limb's muscles remaining in the amputee's stump (which still has normal innervations, and thus is subject to voluntary control) as a control signal to the prosthesis—in other words, it allows amputees to use the same mental process to control their prosthesis as they had used in controlling their physiological parts; however, the technology in that time was not adequate to make clinical application viable. With the development of semiconductor devices technology, and the associated decrease in device size and power requirements, the clinical applications saw promise, and research and development increased dramatically.

DOI: 10.4018/978-1-60960-561-2.ch322

This control approach referred to as myoelectric control has found widespread use for individuals with amputations or congenitally deficient limbs. The methodologies used in this approach of control spanned the range from classical multivariate statistical and syntactic methods, to the newer artificial intelligence (symbolic and connectionist) approaches to pattern processing. In these systems, voluntarily controlled parameters of myoelectric signals from a muscle or muscle group are used to select and modulate a function of a multifunction prosthesis. The essential elements of a myoelectric control of prosthesis devices are shown in the block diagram schematic of Figure 1 (Merletti & Parker, 2004). The Myoelectric control system is based on the noninvasive interfaces designed for casual wear.

In the following sections, the details and related works to the research problem are given; after that, the methodology adopted is explained and discussed, followed by future work, and finalised with conclusion.


BACKGROUND AND RELATED WORK

Continuous myoelectric-controlled devices are one of the challenging research issues, in which the

prosthetic is controlled in a manner proportional to the level of myoelectric activity. Although the success of fitting these systems for single device control is apparent, the extension to control more than one device has been difficult (Hudgins, Parker, & Scott, 1993), but, unfortunately, it is required for those with high-level (above the elbow) limb deficiencies, and the individuals who could stand to benefit from a functional replacement of their limbs (Englehart, Hudgin, & Parker, 2001). It has been proved that the MES signal exhibits a deterministic structure during the initial phase of muscle contraction, as shown in Figure 2 for four types of muscle contractions measured using one surface electrode. Based on this principle, a continuous myoelectric control strategy based on the use of pattern classifiers was the main focus during the last years for most scientists in related fields.

The MES patterns exhibit distinct differences in their temporal waveforms. Within a set of patterns derived from the same contraction, the structure that characterizes the patterns is sufficiently consistent to maintain a visual distinction between different types of contraction. Hudgins et al. (1993), and Lighty, Chappelly, Hudgins, and Englehart (2002) aligned the patterns using a cross-correlation technique, and showed that the ensemble average of patterns within a class pre-

Figure 1. Normal and myoelectric control system



7 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage:

www.igi-global.com/chapter/myoelectric-control-prosthetic-devices-rehabilitation/53631

Related Content

Image Registration Algorithms for Applications in Oncology

Katia Marina Passeraand Luca Tommaso Mainardi (2009). *Handbook of Research on Advanced Techniques in Diagnostic Imaging and Biomedical Applications* (pp. 126-141).

www.irma-international.org/chapter/image-registration-algorithms-applications-oncology/19592

Analysis and Quantification of Motion within the Cardiovascular System: Implications for the Mechanical Strain of Cardiovascular Structures

Spyretta Golemati, John Stoitsisand Konstantina S. Nikita (2009). *Handbook of Research on Advanced Techniques in Diagnostic Imaging and Biomedical Applications* (pp. 34-47).

www.irma-international.org/chapter/analysis-quantification-motion-within-cardiovascular/19586

Internal Radionuclide Dosimetry using Quantitative 3-D Nuclear Medical Imaging

Ioannis Tsougos, George Loudos, Panagiotis Georgoulas, Konstantina S. Nikitaand Kiki Theodorou (2009). *Handbook of Research on Advanced Techniques in Diagnostic Imaging and Biomedical Applications* (pp. 213-228).

www.irma-international.org/chapter/internal-radionuclide-dosimetry-using-quantitative/19597

Assessment of the Truth Foreseen in the Intuition

Carlo Ciulla (2009). *Improved Signal and Image Interpolation in Biomedical Applications: The Case of Magnetic Resonance Imaging (MRI)* (pp. 47-51).

www.irma-international.org/chapter/assessment-truth-foreseen-intuition/22491

Telemedicine Consultations in Daily Clinical Practice

Anton V. Vladzmyrskyy (2011). *Clinical Technologies: Concepts, Methodologies, Tools and Applications* (pp. 1062-1074).

www.irma-international.org/chapter/telemedicine-consultations-daily-clinical-practice/53638