

Chapter 9

Laser Additive Manufacturing of Titanium–Based Implants: A Review

Martin Ruthandi Maina

Jomo Kenyatta University of Agriculture and Technology, Kenya

ABSTRACT

Titanium and its alloys exhibit a unique combination of mechanical, physical properties and corrosion resistance behaviour which makes them desirable for aerospace, industrial, chemical, medical and energy industries. The selective addition of alloying elements to titanium enables a wide range of physical and mechanical properties to be obtained. Ti-based alloys are finding ever-increasing applications in biomaterials due to their excellent mechanical, physical and biological performance. Intense researches are being pursued in the development of new Ti-based alloys with bio-functionalization closer to human bone, owing to their excellent mechanical strength and resilience when compared to alternative biomaterials, such as polymers and ceramics. Several manufacturing techniques are capable of producing porous materials. There is a need to control pore size, shape, orientation and distribution. This work reviews the application of Ti-based alloys in the biomedical industry and also proposes laser additive manufacture process for the manufacture of medical implants.

DOI: 10.4018/978-1-5225-0329-3.ch009

1. INTRODUCTION

Metals and their alloys are widely used as biomedical materials. Metallic biomaterials like stainless steels, Co-Cr alloys, commercially pure titanium and its alloys are extensively used due to their excellent mechanical properties. However, metallic materials sometimes show toxicity and are fractured because of their corrosion and mechanical damages (Yuhua Li. et al, 2014; Kannatey-Asibu, 2009).

The family of titanium alloys offers a wide spectrum of strength and combinations of strength and fracture toughness. This permits optimized alloy selection which can be tailored for a critical component. The development must be performed on the basis of metallurgy and the resultant alloys must have a good balance between mechanical properties and corrosion resistance. Among metallic materials, titanium and its alloys are considered the most suitable materials for biomedical applications due to their superior properties. They satisfy the requirements of implantation materials better than other materials (Elias et al., 2008; Temenoff & Mikos, 2008).

Laser additive manufacturing is one of the processes for manufacture of titanium based alloys. The process has many variables that can influence the soundness and mechanical properties of the resulting part by affecting characteristics such as surface finish, porosity, residual stresses/cracking, microstructure, and texture. The variables include laser type, laser power and power distribution, laser spot size and shape, laser traverse speed, line spacing, layer thickness, deposition pattern, powder shape, size, and size distribution, powder feed rate, powder velocity, substrate temperature, substrate surface finish, substrate thickness, substrate microstructure and texture, and the size and shape of the deposit (Gu et al., 2012; Kobryn & Semiatin, 2001).

The composition of implant biomaterials must be carefully selected to avoid adverse reactions. Metals such as Ti, Zr, Nb, Mo, Ta, Sn are non-toxic and have good compatibility. β -type Ti-based alloys have high strength and low elastic modulus compared to pure titanium, hence they are used as a starting material for the improvement of mechanical properties of porous compacts. Although, fabrication of implants from materials with lower elastic modulus can reduce stress shielding effect, the modulus mismatch to bone is still substantial. To provide a way for living bone to attach itself permanently to an implant, an artificial bone should have a porous structure. Porous titanium implants demonstrate an important gain in promoting tissue in-growth and in the firm securing of an implant (Elias et al., 2008).

Porous materials in implants are increasingly attracting widespread interest of researchers and desirable by the biomedical industries. The mechanical properties of porous Ti-based alloys are dependent on porosity, pore morphology, pore size distribution and microstructure. There is a need to come up with a novel method of manufacture of the porous Ti based alloys with a keen control of pore size, shape, orientation and distribution. There is also need to develop analytical models to predict

10 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage: www.igi-global.com/chapter/laser-additive-manufacturing-of-titanium-based-implants/149843

Related Content

Morphology and Functionalization of Metal Foils and Other Surfaces for Electrochemical Applications

Martin Rozman and Miha Lukšič (2022). *Handbook of Research on Tribology in Coatings and Surface Treatment* (pp. 359-389).

www.irma-international.org/chapter/morphology-and-functionalization-of-metal-foils-and-other-surfaces-for-electrochemical-applications/301925

Advanced Machining Techniques for Fiber-Reinforced Polymer Composites

Inderdeep Singh and Kishore Debnath (2017). *Materials Science and Engineering: Concepts, Methodologies, Tools, and Applications* (pp. 112-135).

www.irma-international.org/chapter/advanced-machining-techniques-for-fiber-reinforced-polymer-composites/175691

Development and Assessment of a New Flow-Through Test Instrument to Study Wear and Erosion Effects of Nanofluids

Gustavo J. Molina, Fnu Aktaruzzaman, Valentin Soloiu, Mosfequr Rahman and Kenshantis Martin (2017). *International Journal of Surface Engineering and Interdisciplinary Materials Science* (pp. 58-74).

www.irma-international.org/article/development-and-assessment-of-a-new-flow-through-test-instrument-to-study-wear-and-erosion-effects-of-nanofluids/173733

3D Printing and Actor-Network Theory

Graham Harman (2017). *3D Printing: Breakthroughs in Research and Practice* (pp. 48-56).

www.irma-international.org/chapter/3d-printing-and-actor-network-theory/168213

Wear and Friction Effects on Engine Materials of N-Butanol in Ultra-Low Sulfure Diesel Mixtures

Gustavo J. Molina, John L. Morrison, Valentin Soloiu and Cesar Carapia (2022). *International Journal of Surface Engineering and Interdisciplinary Materials Science* (pp. 1-11).

www.irma-international.org/article/wear-and-friction-effects-on-engine-materials-of-n-butanol-in-ultra-low-sulfure-diesel-mixtures/298707