Chapter 45 A Novel Delay–Based GFSK Demodulator in 65 nm CMOS for Low Power Biomedical Applications

Meng Fu

University of Melbourne, Australia

Stan Skafidas University of Melbourne, Australia

Iven Mareels University of Melbourne, Australia

ABSTRACT

This article describes how, in recent years, with the development of microelectronics, implantable electronic devices have been playing a significant role in modem medicine. Examples of such electronic implant devices are, for instance, retinal prosthesis and brain implants. It brings great challenges in low power radio frequency (RF) and analog designs. This article presents a low power Gaussian frequency shift keying (GFSK) demodulator designed for Medical Implant Communications Service (MICS) band Receiver. This demodulator utilizes a novel structure that a wide IF range can be handled and presents the smallest $\Delta f/f$ ratio in any published GFSK demodulators. In theory the demodulation method can be applied to any RF frequency. The demodulator draws 550uA from a 1 V power supply. A maximum data rate of 400 Kbits/s can be achieved within the 300 KHz channel bandwidth defined by MICS. A simulated signal-to-noise ratio (SNR) of 15.2dB at AWGN channel is obtained to achieve 10^3 bit error rate (BER). This demodulator is fabricated on 65-nm CMOS and occupies 0.12mm2 silicon area.

DOI: 10.4018/978-1-7998-1204-3.ch045

INTRODUCTION

The MICS band is an ultra-low power, unlicensed mobile radio service for implanted medical device. Ten 300 KHz channels are allocated at 402-405 MHz frequency band ("FCC Rules and Regulations," 2002). Implanted wireless device has a requirement of high reliability and ultra-low power consumption. Recent CMOS technology has enabled the design of reliable, low power and cost effective wireless transceiver. This demodulator is designed on 65nm technology under 1 V supply. It intends to be used for the epi-retinal prosthesis (Bionic Eye) to demodulate data for a 32x32 electrode array. In order to support these 1024 electrodes, a data rate more than 300 Kbps is required to make fully functional. Achieving high date rate and low power consumption from this limited 300 KHz bandwidth are challenges in the design of such system. So far, amplitude shift keying (ASK) and binary shift keying (BFSK) have been widely used in biomedical implants (Ghovanloo, Beach, Wise, & Najafi, 2002; Ghovanloo & Najafi, 2003). However, these two methods are not bandwidth efficient. This paper presents a demodulator that can work for both continuous phase BFSK and GFSK modulation schemes. Using GFSK, a data rate up to 400 Kbit/s is achieved within the given 300 KHz bandwidth. This demodulator is also immune to a frequency offset up to 4 MHz. A BER of 10⁻³ is obtained at SNR of 15.2 dB at the input.

DEMODULATOR ARCHITECTURE

The Architecture of the RF receiver is shown in Figure 1. The RF frequency is down converted from MICS band, 402-405MHz to around 5 MHz. A channel-select filter is used to reject the image and select the desired channel. Then the limiter amplifies the IF signal to rail-to-rail swing. Block diagram of proposed GFSK demodulator is shown in the dashed box of the figure. It consists of four blocks: delay line, delay flip-flop (DFF), gm-c low pass filter, and decision-making circuit. The delay element delays the IF input by approximate half the bit period. Then the DFF uses the delayed signal as the clock to sample the original signal. So the frequency information of GFSK signal is mapped into the density and width



Figure 1. Block diagram of receiver with the proposed GFSK demodulator

11 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage:

www.igi-global.com/chapter/a-novel-delay-based-gfsk-demodulator-in-65-nm-

cmos-for-low-power-biomedical-applications/243147

Related Content

Detection of Anomalous Transactions in Mobile Payment Systems

Ibrar Hussainand Muhammad Asif (2020). *International Journal of Data Analytics (pp. 58-66).* www.irma-international.org/article/detection-of-anomalous-transactions-in-mobile-payment-systems/258921

Proximate Breast Cancer Factors Using Data Mining Classification Techniques

Alice Constance Mensahand Isaac Ofori Asare (2019). International Journal of Big Data and Analytics in Healthcare (pp. 47-56).

www.irma-international.org/article/proximate-breast-cancer-factors-using-data-mining-classification-techniques/232335

Railway Operations Models: The OR Approach

Sundaravalli Narayanaswami (2018). Intelligent Transportation and Planning: Breakthroughs in Research and Practice (pp. 745-767).

www.irma-international.org/chapter/railway-operations-models/197160

Improvement in Task Scheduling Capabilities for SaaS Cloud Deployments Using Intelligent Schedulers

Supriya Sawwashere (2021). International Journal of Big Data and Analytics in Healthcare (pp. 1-12). www.irma-international.org/article/improvement-in-task-scheduling-capabilities-for-saas-cloud-deployments-usingintelligent-schedulers/287104

Land Cover Classification Using the Proposed Texture Model and Fuzzy k-NN Classifier

Jenicka S. (2018). *Optimization Techniques for Problem Solving in Uncertainty (pp. 226-261).* www.irma-international.org/chapter/land-cover-classification-using-the-proposed-texture-model-and-fuzzy-k-nnclassifier/206636