Chapter 2 The Limitations of Static Traditional Occlusal Indicators Compared to the Advantages of Quantifiable Dynamic Occlusal Indicators

Sarah Qadeer b https://orcid.org/0000-0002-1389-5996 Dr. Med. Dent. University of Basel, Switzerland

ABSTRACT

The traditional occlusal indicators used in dental practice are articulation papers, Shim-stock foils, elastomeric impression materials, and occlusal wax strips. These static dental materials have been widely believed to have occlusal force descriptive capability. However, modern material studies are challenging the widespread belief that occlusal indicator materials can measure differing occlusal force levels. Chapter two evaluates the force reporting limitations of these static occlusal indicators, and discusses how clinicians subjectively interpret their appearance characteristics to determine differing occlusal force levels. This chapter then compares these non-digital occlusal indicators to the T-Scan computerized occlusal analysis technology, that records and displays precise, quantifiable, relative occlusal force variances, and occlusal contact timing sequences. This digital data aids the clinician in making a more accurate occlusal analysis, and can guide the clinician in the correction of occlusal contact force and timing abnormalities, thereby eliminating the subjectivity that is inherent with traditional occlusal indicator use. To further support the superiority of digital occlusal analysis with the T-Scan technology, a systematic review that compared the many static, non-digital occlusal indicators to quantifiable dynamic digital occlusal indicators is included, as is an articulating paper/foil/T-Scan comparative analysis. A recently published systematic review that evaluated occlusal contacts during Maximum Intercuspation with differing occlusal indicators (both traditional and digital) is also described. Next, this chapter details the diagnostic occlusal capabilities of the T-Scan's digital force and timing data by presenting 2 separate studies that compared measured closure and excursive occlusal contact force and timing parameters in orthodontic and non-orthodontic young adults. A final commentary describes the clinical pitfalls of using maximally invasive Subjective Interpretation to choose occlusal contacts for

DOI: 10.4018/978-1-6684-9313-7.ch002

adjustment, instead of employing minimally invasive T-Scan-guided occlusal contact selection. This last section clearly outlines to the reader that both patients and dentists markedly benefit from implementing the T-Scan 10 occlusal measurement technology.

INTRODUCTION

Occlusal indicators are used in prosthetic dentistry to determine occlusal contact locations, and for the recording and transferring of patient's interocclusal relationship onto an articulator for the fabrication of dental prostheses. The conventional or traditional methods used in clinical practice for occlusal contact selection during occlusal adjustment procedures are non-digital. Static dental materials are placed between opposing teeth to imprint, or mark with color, the occlusal contacts. Commonly employed non-digital occlusal indicators are often combined with the patient's subjective occlusal "*feel*" verbal feedback, to help guide the clinician in occlusal contact forcefulness detection.

These static materials display a wide range of physical properties (viscosity, elasticity, volumetric shrinkage, distortion, and crumpling), which contribute to their clinical use inaccuracies.

In clinical practice when making occlusal adjustments, the most commonly used non-digital occlusal indicators are:

- Articulation paper strips that leave ink marks on the teeth where occlusal contacts exist
- **Shim-stock foils,** which are tugged and pulled from between the teeth, to detect withdrawal resistance that supposedly indicates the presence of forceful tooth contacts.
- Elastomeric impression materials, which when injected between opposing teeth to locate occlusal contacts, are displaced completely where there is tooth contact.
- **Occlusal wax sheets,** which are softened and then imprinted by opposing teeth. Wax perforations or apparent wax thinness indicate occlusal contact, or near contact

The properties of an 'ideal' interocclusal recording material are (Malone & Koth, 1989):

- Limited initial resistance during closure, to avoid displacement of both mobile teeth and the mandible, itself
- Excellent dimensional stability after setting
- High resistance to compression after polymerization
- Ease of manipulation
- The absence of any adverse effects on the tissues involved in the recording procedures
- Accurately records the incisal or occlusal surfaces of teeth
- Ease of record verification

There is no ideal occlusal recording material available to clinicians that possess all of these properties (Malone & Koth, 1989). The static materials listed above all demonstrate varying degrees of limitation, when a clinician is choosing which contact(s) appear to be forceful during an occlusal adjustment procedure.

78 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage: <u>www.igi-global.com/chapter/the-limitations-of-static-traditional-occlusal-</u> <u>indicators-compared-to-the-advantages-of-quantifiable-dynamic-occlusal-</u> <u>indicators/363263</u>

Related Content

Joint Vibration Analysis (JVA) and the Diagnostic Process in Temporomandibular Disorders (TMD)

Ray M. Becker (2025). Handbook of Research on T-Scan Technology Applications in Dental Medicine (pp. 875-976).

www.irma-international.org/chapter/joint-vibration-analysis-jva-and-the-diagnostic-process-in-temporomandibulardisorders-tmd/363269

Recent Trends in 3D Printing of Dental Models: Rapid Prototyping in Dental Implants

Kayalvizhi Mohan (2019). *Computational Techniques for Dental Image Analysis (pp. 217-237).* www.irma-international.org/chapter/recent-trends-in-3d-printing-of-dental-models/216051

Airway Changes Detected by Cone Beam Computed Tomography (CBCT) in Muscular

Temporomandibular Joint Dysfunction Subjects Treated With Disclusion Time Reduction (DTR) Atul P. Sattur (2025). Handbook of Research on T-Scan Technology Applications in Dental Medicine (pp. 1431-1546).

www.irma-international.org/chapter/airway-changes-detected-by-cone-beam-computed-tomography-cbct-in-musculartemporomandibular-joint-dysfunction-subjects-treated-with-disclusion-time-reduction-dtr/363273

Thermal Evaluation of Myogenous Temporomandibular Disorders and Myofascial Trigger Points in the Masticatory Muscles

Denise Sabbagh Haddad, Marcos Leal Brioschiand Emiko Saito Arita (2017). Oral Healthcare and Technologies: Breakthroughs in Research and Practice (pp. 329-350).

www.irma-international.org/chapter/thermal-evaluation-of-myogenous-temporomandibular-disorders-and-myofascialtrigger-points-in-the-masticatory-muscles/178989

Image Segmentation Using Contour Models: Dental X-Ray Image Segmentation and Analysis

Kavitha G., Muthulakshmi M.and Latha M. (2019). *Computational Techniques for Dental Image Analysis* (pp. 62-85).

www.irma-international.org/chapter/image-segmentation-using-contour-models/216043