

Chapter 8

Biosensor Based on Giant Magnetoresistance Material

Mitra Djamal

Institut Teknologi Bandung, Indonesia

Ramli

Institut Teknologi Bandung & Universitas Negeri Padang, Indonesia

Yulkifi

Institut Teknologi Bandung & Universitas Negeri Padang, Indonesia

Suprijadi

Institut Teknologi Bandung, Indonesia

Khairurrijal

Institut Teknologi Bandung, Indonesia

ABSTRACT

In recent years, giant magnetoresistance (GMR) sensors have shown a great potential as sensing elements for biomolecule detection. The resistance of a GMR sensor changes with the magnetic field applied to the sensor, so that a magnetically labeled biomolecule can induce a signal. Compared with the traditional optical detection that is widely used in biomedicine, GMR sensors are more sensitive, portable, and give a fully electronic readout. In addition, GMR sensors are inexpensive and the fabrication is compatible with the current VLSI (Very Large Scale Integration) technology. In this regard, GMR sensors can be easily integrated with electronics and microfluidics to detect many different analytes on a single chip. In this article, the authors demonstrate a comprehensive review on a novel approach in biosensors based on GMR material.

INTRODUCTION

The discovery of biosensors began in the early 1960s when Clark and Lyons first reported on amperometric glucose detection using the enzyme glucose oxidase (Clark & Lyons, 1962). Later, biosensors found a wide area of applications in industry, clinical diagnostics, and ecology. At the same time, biosensor science became an independent area of modern analytical chemistry involving approaches of biochemistry and physical chemistry. A biosensor is generally defined as an analytical device, which makes use of a biological molecular recognition component connected to a transducer to generate a quantifiable electronic output signal, in response to a biological or chemical analyte (Li, Sun, Wilson, White, Pourmand, & Wang, 2006).

Biosensors promise to provide analytically powerful and inexpensive alternatives to standard bacteriological methods used today for bacteria detection. Biosensors are analytical devices that respond selectively to analytes or bacteria by use of a combination of a biological recognition system and a physical transducing element. The biological response is converted by the transducer to an electrical or optical signal that is directly related to the concentration of the analyte of interest or bacteria. A biosensor's recognition system may consist of one of many biological components including: enzymes, antigens, antibodies, nucleic acids, cells, cell organelles or receptor molecules. All of these biological components are capable of interacting very specifically with an analyte of interest or with bacteria and to give a biosensor its unique specificity.

Nowadays, accurate, rapid, cheap and selective analysis is required for clinical and industrial laboratories. Magnetoresistive biosensors seem to be among the best candidates to meet these criteria. Since the late 1990s, magnetoelectronics (Xu, Yu, Michael, Han, Osterfeld, White, Pourmand, & Wang, 2008) has emerged as one of several new platform technologies for biosensor and

biochip development. This technology is based on the detection of biologically functionalized micrometer or nanometer-sized magnetic labels, using high-sensitivity microfabricated magnetic-field sensors.

Another concept for DNA detection was introduced by Baselt, Lee, Natesan, Metzger, Sheehan, and Colton (1998). It is based on detecting the stray field of magnetic labels by embedded magnetoresistive sensors. Such a technique, apart from being sensitive, flexible and compatible with standard CMOS fabrication, also has the advantage of directly providing an electronic signal that is suitable for automated on-chip analysis. Furthermore, the possibility to attract the magnetic labels by suitable gradient fields also opens up a completely new prospect that is not possible in such a way for any of the competing detection schemes, i.e., the selective on-chip manipulation of desired biomolecules (Leech, 1994). This option could be used, for example, to reduce incubation times during hybridization, or to test binding forces of adhered molecules (Prinz, 1998). Due to these opportunities, magnetoresistive biosensors represent a very promising alternative detection unit for lab-on-a-chip devices, and since the first publication in 1998, a number of research groups initiated work on that subject (Baselt, Lee, Natesan, Metzger, Sheehan, & Colton, 1998).

A group at the Naval Research Laboratory (NRL) and NVE Corporation in the US first demonstrated a magnetic biosensor system which they called BARC (Rife, Miller, Sheehan, Tamanaha, Tondra, & Whitman, 2003). Their detector employed a giant magnetoresistive multilayer stack that is less sensitive than properly designed spin valve (SV) or magnetic tunnel junction (MTJ) detectors (Wang & Taratorin, 1999). Another group in Germany demonstrated that giant magnetoresistive (GMR) multilayer biodetection was superior to fluorescent biodetection (Schotter, Kamp, Becker, Puhler, Reiss, & Brückl, 2004). Groups in Portugal and Netherlands have deployed SV sensors coupled with coils at proximity for

14 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage:

www.igi-global.com/chapter/biosensor-based-giant-magnetoresistance-material/65706

Related Content

Agent-Based Patient Scheduling

Einar Jakob Hovland and Even Åby Larsen (2008). *Encyclopedia of Healthcare Information Systems* (pp. 46-53).

www.irma-international.org/chapter/agent-based-patient-scheduling/12921

The Creation of Users: A Learning Experience in Information System Development

Jan Aidemark and Linda Askenäs (2018). *International Journal of E-Health and Medical Communications* (pp. 74-88).

www.irma-international.org/article/the-creation-of-users/201549

A Web Based Software System for Bone and Joint Infections in Children

Vaia Tsioutsou, Vasileios Syrimpeis, Giannis Tzimas, George Sdougkos, Athanasios Tsakalidis and Elias Panagiotopoulos (2012). *International Journal of User-Driven Healthcare* (pp. 44-56).

www.irma-international.org/article/web-based-software-system-bone/75181

Fuzzy Logic-Based Predictive Model for the Risk of Type 2 Diabetes Mellitus

Peter Adebayo Idowu and Jeremiah Ademola Balogun (2019). *International Journal of E-Health and Medical Communications* (pp. 56-78).

www.irma-international.org/article/fuzzy-logic-based-predictive-model-for-the-risk-of-type-2-diabetes-mellitus/227697

Transition to ISO 15189 : 2012 for Cytopathology Laboratories Part 1: General and Management Requirements

Eleftherios Vavoulidis, Stavros Archondakis, Maria Nasioutziki, Ourania Oustambasidou, Angelos Daniilidis, Konstantinos Dinas and Aristotelis Loufopoulos (2016). *International Journal of Reliable and Quality E-Healthcare* (pp. 1-21).

www.irma-international.org/article/transition-to-iso-15189--2012-for-cytopathology-laboratories-part-1/159067